

Will Acoustic Gas Thermometry Lead to a New Temperature Scale?

L. Pitre^S

Institut National de Métrologie, Laboratoire National de Métrologie et d'Essai, CNAM, Saint Denis, France

M.R. Moldover^C and W. Tew

Process Measurements Division, National Institute of Standards and Technology, Gaithersburg, MD, USA
michael.moldover@nist.gov

We used a quasi-spherical cavity as an acoustic and microwave resonator to measure the thermodynamic temperatures, T , of the triple points of mercury, argon, neon, and equilibrium hydrogen, and to measure the difference $T - T_{90}$, in the range 7 to 273 K (T_{90} is the temperature on the International Temperature Scale of 1990 ITS-90). In the range 7 to 24.5 K, we achieve uncertainties that are comparable to, or smaller than, those achievable using the interpolating constant volume gas thermometer as currently defined on the ITS-90. In the range 90 to 273 K, the present results for $T - T_{90}$ obtained using a helium-filled, copper-walled, quasi-spherical cavity agree with earlier results obtained using argon-filled, steel-walled, or aluminum-walled, spherical cavities. The agreement confirms our understanding of both acoustic and microwave cavity resonators, and demonstrates that resonators function as primary thermometers spanning wide temperature ranges. The mutually consistent acoustic thermometry data from several laboratories imply that the values of $(T - T_{90})/T_{90}$ are 5 times larger than the uncertainty of T/T_{90} near 150 and 400 K. They also imply that the derivative dT/dT_{90} is too large by approximately 10^{-4} near 273.16 K and that dT/dT_{90} has a discontinuity of $4 \cdot 10^{-5}$ at 273.16 K. Thus, a new temperature scale based on acoustic gas thermometry could be significantly more accurate than ITS-90.